

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606800
Email: info@kemsa.co.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA/PROC/OIT02/2024-2026

Date: 14th October, 2024

CLARIFICATION NO. 1

To: All bidders,

TENDER NO: KEMSA/OIT02/2024-2026 - SUPPLY OF LABORATORY REAGENTS AND CONSUMABLES

The following queries were raised by an interested bidders and we hereby respond as follows;

QUESTION 1

Item no. 26 - VDRL Test Strips - Syphilis Screening:

Can one quote for syphilis test strips and if so, can one quote a pack of 30s?

Whereas the description for item no. 26 in the tender document is VDRL Test Strips, the detailed product specifications are for VDRL Reagent. Kindly clarify weather to quote for both VDRL STRIPS and VDRL Reagents or either.

For Item 26, could you confirm if you are looking for test strips or a test reagent? The short description specifies "VDRL TEST STRIPS - SYPHILIS SCREENING," whereas the long description refers to it as "VDRL REAGENT."

RESPONSE

We require VDRL Test Strips - Syphilis Screening. Kindly find revised item specifications attached.

- The pack should be for 50 tests

QUESTION 2

For Item No. 7 Brucella Abortus Kit (Antigen Test) and Item No. 8 Brucella Melitensis Kit (Antigen Test);

Can one quote 5ml which is an equivalent to 100 tests.

RESPONSE

Yes - as long as it can run 100 tests

QUESTION 3

For Item No. 14- Hepatitis B (Hbsag) Test Strips, can one quote a pack of 30s/kit which will result to 4334 kits?

RESPONSE

Kindly stick to tender requirements

QUESTION 4

Can we quote and submit the items below in the following pack sizes?

- Item No. 1 Asot (Pack of 100 tests)
- Item No. 24 Rheumatoid factor (100 tests)

RESPONSE

Kindly stick to tender requirements

QUESTION 5

For Item 6, could you please clarify whether you need IgM only or the IgG/IgM blend? The short description indicates "ANTI D IgM Monoclonal," while the detailed specifications mention "ANTI D IgG/IgM blend reagent."

RESPONSE

We need ANTI D IgM Monoclonal



QUESTION 6

For Item 27, could you provide the brand and model of the urine machine you are using? This will help us ensure that the urine strips we quote are compatible with your equipment.

RESPONSE

We don't procure for specific Urine machines; we require strips for Physical visualization using the colour codes on the tin (Package) holding the strips

QUESTION 7


Item No. 13 NL05TES094: Helicobacter Pylori Rapid Kit (Antigen)

The unit pack in the price schedule is pack of 20's, while the technical specification says pack of 25's. Please confirm the correct pack size.

RESPONSE

The correct pack size is Kit X 20 Tests

Yours faithfully,



Moses Sudi

Procurement Director

FOR: CHIEF EXECUTIVE OFFICER

Revised item specifications

Item no. 26 - SYPHILIS RAPID TEST KIT

Intended use: Syphilis Whole Blood/Serum/Plasma Test strip is a rapid test for the visual detection of antibodies of Treponema Pallidum (TP) in whole blood/serum/plasma as an aid in the diagnosis of syphilis infection.

Test Principle: lateral flow chromatographic immunoassay

λIndividually sealed foil pouches containing:

λOne dip strip device

λOne desiccant

λPackage insert (instruction for use)

λISO/CE standards - Should meet the requirement of ISO:13485/ CE Mark/WHO prequalified

λMust be registered by PPB

λLabeled "GOK KEMSA" using indelible ink upon award of contract

SUBMIT A PACK FOR EVALUATION

