

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsa.co.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA /ONT03/2021-2023

Date: 23rd July, 2021

CLARIFICATION 1

To: All bidders,

Dear Sir/Madam,

REF: TENDER KEMSA/ONT03/2021-2023 - SUPPLY OF HEALTH TECHNOLOGIES (SURGICAL DRESSINGS, RESUCSITAIRE, CATHETERS, CORD CLAMP, SURGICAL GLOVES, CANNULAS, ADMINISTRATION SETS & SYRINGES)

The following question was raised by an interested bidder and response provided:

Question:

We are seeking clarification on the Gloves water weight must it accommodate 1ltr or it can be less?

RESPONSE:

The examination gloves latex medium size should accommodate 800ml to 1000ml of water.

Yours faithfully

A handwritten signature in black ink, appearing to be 'Edward Buluma', written over a horizontal line.

EDWARD BULUMA
FOR: AG. CHIEF EXECUTIVE OFFICER