

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsa.co.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA/ GF ATM TB - OIT 001 / 2021-22

Date: 27th September, 2021

ADDENDUM 1:

Attn.: All Prospective Bidders

RE: GF ATM TB COVID-19 RESPONSE 2021/2022-OIT -001 - SUPPLY OF LABORATORY EQUIPMENT

In accordance with ITT Clause 27.1 (Clarification of Tenders) of the Open International Tender (OIT) issued under the above reference, we hereby respond to questions raised by a prospective bidder as follows:

Question 1:

With reference to the subject, please clarify on the Dimensions for Biosafety Cabinets, Item No 2. Should it be in CM or MM?

On the Technical Specification, It is indicated as "Dimensions Exterior in cm (L X W X H): L (min 500-max 900) x W (min 1000- max1900) x H (min 1200-max 1550).

Response 2:

The unit of Measure is in MM.

Yours faithfully,

A handwritten signature in blue ink, appearing to read "Edward Buluma", is written over a blue circular stamp.

EDWARD BULUMA
FOR: Ag. CHIEF EXECUTIVE OFFICER