

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsaco.ke

All Correspondence should be addressed to Chief Executive Officer.

When replying please quote our ref:

KEMSA/WB- CERP/- RFQ 01 /22

Date: 28th April 2022

CLARIFICATION 1:

Attn.: All prospective bidders

RE: TENDER NO: KEMSA/WB-CERP/RFQ 01/2021-2022 SUPPLY, INSTALLATION AND COMMISSIONING OF ASSORTED MEDICAL EQUIPMENT.

The following is a query raised by an interested bidder and responses made as below:

Question 1:

Lot 1: Item 1: Patient Monitors

According to technical specifications number 3.2, the display size of the monitor is indicated as 21 inch. The size of the monitor seems is too big for a patient monitor. Is 15 inch acceptable?

Response:

The monitor screen size should be 21 inch.

Bidders should note as below;

Lot 1: Item 1: Patient Monitors

The certificate of quality as indicated on page 9 of the tender document should be, ISO 13485:2016 and not ISO 13485:2003

As per specifications 6.2 internal rechargeable battery; Maintenance free type, should be up to 3 hours operating time and not 8 hours.

All other terms and conditions remain.

Yours faithfully,

DR. SILAS E. NJERU (PhD)
AG. DIRECTOR PROCUREMENT
FOR: AG. CHIEF EXECUTIVE OFFICER