

# KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area  
P.O. Box 47715, 00100 GPO,  
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400  
GSM +254 719 033000, +254 733 606600  
Email: info@kemsaco.ke

*All Correspondence should be addressed to Chief Executive Officer*

When replying please quote our ref:

KEMSA /ONT16/2021-2023

Date: 22nd April, 2022

## CLARIFICATION 1

To: All bidders,

Dear Sir/Madam,

**REF: TENDER NO: KEMSA/ONT16/2021-2023: - FOR SUPPLY OF ASSORTED BLOOD BAGS**

The following questions were raised by interested bidders and responses provided:

### **QUESTION 1:**

Regarding to the specification, Design and shape: it is written "Slit on both sides of the bags should be enough to accommodate 5-10ml test tube, please clarify since we know the function of this slit should be used for fold the blood bags before putting them in centrifuge machines instead of hold test tubes.

### **RESPONSE**

Regarding design and shape the SLIT has many purposes, that's why we said TO FIT 5-10 mls specimen bottles. Therefore, it remains the way it is.

### **QUESTION 2:**

For solutions of triple and quadruple, please advise if you need CPDA-1 or CPD+SAGM.

Kindly clarify on the attached specification of Quadruple blood bags. It requires CPDA-1 +addictive solution as per the bid document. Should it be CPD+SAGM or CPDA-1+SAGM?.

### **RESPONSE**

Triple and quadruple bags anticoagulant is CPD + Additive solution (100mls for 450ml blood bag).

### **QUESTION 3:**

How many pcs we should put in foil bag and the samples quantity required for submission

**RESPONSE**

On packaging

- (1). Single 5-7 units in a foil
  - (2). Double, Triple, Quadruple 3-5 Units in a foil for Sample Specimens
- 2 (Two) Loose units and 1(One) foil for all i.e. (Single, double, Triple and Quadruple)

Yours faithfully



**DR. SILAS E. NJERU (PhD)**  
**AG. DIRECTOR PROCUREMENT**  
**FOR: AG. CHIEF EXECUTIVE OFFICER**