

KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area
P.O. Box 47715, 00100 GPO,
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400
GSM +254 719 033000, +254 733 606600
Email: info@kemsa.co.ke

All Correspondence should be addressed to Chief Executive Officer

When replying please quote our ref:

KEMSA /OIT21/2021-2023

Date: 31st May, 2022

CLARIFICATION 2

To: All bidders,

Dear Sir/Madam,

RE: TENDER NO: KEMSA/OIT21/2021-2023 - SUPPLY OF LABORATORY CONSUMABLES

The following questions were raised by interested bidders and responses provided:

QUESTION 1:

Kindly also clarify on item no. 7 retractable lancet. When you state approximately 28Gauge - is it ok to offer less than 28G or more than 28G or it must be 28G?

RESPONSE

The Needle Gauge should be 28.

QUESTION 2:

Kindly clarify if the Brucella test kits are serology kits or cassettes.

RESPONSE

Refer to clarification 1 dated 27th May, 2022 which is available on KEMSA website and PPIP.

QUESTION 3:

Please also confirm, if we are giving a sample of HBA1C Analyzer?

RESPONSE

Yes, sample is required for the item.

Yours faithfully

DR. SILAS E. NJERU (PhD)
AG. DIRECTOR PROCUREMENT
FOR: CHIEF EXECUTIVE OFFICER