

# KENYA MEDICAL SUPPLIES AUTHORITY

Commercial Street, Industrial Area  
P.O. Box 47715, 00100 GPO,  
Nairobi, Kenya



Tel: +254 20 3922000, Fax: +254 203922400  
GSM +254 719 033000, +254 733 606600  
Email: info@kemsaco.ke

*All Correspondence should be addressed to Chief Executive Officer*

When replying please quote our ref:

KEMSA/PROC/ONT18/2023

Date: 29<sup>th</sup> March, 2023

## CLARIFICATION 2

To: All bidders,

Dear Sir/Madam,

**REF: TENDER NO: KEMSA/ONT18/2022-2024:-SUPPLY OF NON-PHARMACEUTICALS (II)**

The following question was raised by interested bidders and responses provided:

### **QUESTION 1:**

As for the Tender No: KEMSA/ONT18/2022-2024, Item 8. Under Water Seal drainage bottle, the capacity on the specifications is marked as 2000, 2500ML please confirm the correct size. Also confirm if it is okay to provide tube that is more than 80cm, like 92cm.

### **RESPONSE 1:**

- The capacity is between 2000 to 2500mls.
- A tube of 80 to 95cm is allowed.

### **QUESTION 2:**

We hereby wish to seek clarification on the Below specification for Item No 21. Gloves, Surgical Size 6. Product parameters: Latex material Pre- powdered: The gloves to have **straight fingers**. Are we eligible to quote for **curved fingers** since straight fingers specification is for EXAMINATION GLOVES while curved is for Surgical gloves?

### **RESPONSE 2:**

Please stick to product specifications.

Yours faithfully

A handwritten signature in blue ink, appearing to be 'J. Kinoti', written over a circular stamp.

**Dr. Justus Kinoti (Ph.D)**  
**Ag. DIRECTOR, PROCUREMENT**  
**FOR: CHIEF EXECUTIVE OFFICER**